



Scope of Accreditation for Calibration

As per ISO/IEC 17025:2017

Page No. 1 of 6

CAB Name : Tycon Engineering

Address : Plot no. 361,362, samaypur colony near C.I.A office sector-56 Ballabhgarh
Faridabad-121004, Haryana

Certificate No: C-0006

Issue date : 16.01.2025

Validity : 15.01.2029

Amendment date : N/A

Mechanical Calibration

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	CMC(±)	Remarks/Method used
Facility - Permanent				
01	weight	1mg, E2 Accuracy class and Coarser	0.0013 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
02	weight	2mg, E2 Accuracy class and Coarser	0.0013 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
03	weight	5mg, E2 Accuracy class and Coarser	0.0013 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
04	weight	10mg, E2 Accuracy class and Coarser	0.0013 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
05	weight	20mg, E2 Accuracy class and Coarser	0.0013 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
06	weight	50mg, E2	0.0013 mg	Using E1 Accuracy class

Ratan Singh Rana
Director, IQAS



Scope of Accreditation for Calibration

As per ISO/IEC 17025:2017

Page No. 2 of 6

CAB Name : Tycon Engineering

Address : Plot no. 361,362, samaypur colony near C.I.A office sector-56 Ballabhgarh
Faridabad-121004, Haryana

Certificate No: C-0006

Issue date : 16.01.2025

Validity : 15.01.2029

Amendment date : N/A

Mechanical Calibration

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	CMC(±)	Remarks/Method used
		Accuracy class and Coarser		standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
07	weight	100mg, E2 Accuracy class and Coarser	0.0022 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
08	weight	200mg, E2 Accuracy class and Coarser	0.0022 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
09	weight	500mg, E2 Accuracy class and Coarser	0.0022mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
10	weight	1g, E2 Accuracy class and Coarser	0.0030 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
11	weight	2g, E2 Accuracy class and Coarser	0.0030 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity

Ratan Singh Rana
Director, IQAS



Scope of Accreditation for Calibration

As per ISO/IEC 17025:2017

Page No. 3 of 6

CAB Name : Tycon Engineering

Address : Plot no. 361,362, samaypur colony near C.I.A office sector-56 Ballabhgarh
Faridabad-121004, Haryana

Certificate No: C-0006

Issue date : 16.01.2025

Validity : 15.01.2029

Amendment date : N/A

Mechanical Calibration

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	CMC(±)	Remarks/Method used
				5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
12	weight	5g, E2 Accuracy class and Coarser	0.0050 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 5 g and d: 0.001 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
13	weight	10g, E2 Accuracy class and Coarser	0.010 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 220 g and d: 0.01 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
14	weight	20g, E2 Accuracy class and Coarser	0.010 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 220 g and d: 0.01 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
15	weight	50g, E2 Accuracy class and Coarser	0.010 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 220 g and d: 0.01 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
16	weight	100g, E2 Accuracy class and Coarser	0.020 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 220 g and d: 0.01 mg, Substitution Method by ABBA

Ratan Singh Rana
Director, IQAS



Scope of Accreditation for Calibration

As per ISO/IEC 17025:2017

Page No. 4 of 6

CAB Name : Tycon Engineering

Address : Plot no. 361,362, samaypur colony near C.I.A office sector-56 Ballabhgarh
Faridabad-121004, Haryana

Certificate No: C-0006

Issue date : 16.01.2025

Validity : 15.01.2029

Amendment date : N/A

Mechanical Calibration

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	CMC(±)	Remarks/Method used
				Cycles as per OIML R-111:2004
17	weight	200g, E2 Accuracy class and Coarser	0.030 mg	Using E1 Accuracy class standard weights 1 mg to 200 g and Weighing Balance Capacity 220 g and d: 0.01 mg, Substitution Method by ABBA Cycles as per OIML R-111:2004
18	weight	500g, E2 Accuracy class and Coarser	0.080 mg	Using E1 Accuracy class standard weight 500 g and Weighing Balance Capacity 1 kg and d: 0.0001 g, Substitution Method by ABBA Cycles as per OIML R-111:2004
19	weight	1kg, E2 Accuracy class and Coarser	0.20 mg	Using E1 Accuracy class standard weight 1kg and Weighing Balance Capacity 1 kg and d: 0.0001 g, Substitution Method by ABBA Cycles as per OIML R-111:2004
20	weight	2kg, E2 Accuracy class and Coarser	1.0 mg	Using E1 Accuracy class standard weight 2 kg and Weighing Balance Capacity 10 kg and d: 0.001 g, Substitution Method by ABBA Cycles as per OIML R-111:2004
21	weight	5kg, E2 Accuracy class and Coarser	1.0 mg	Using E1 Accuracy class standard weight 5 kg and Weighing Balance Capacity 10 kg and d: 0.001 g, Substitution Method by ABBA Cycles as per OIML R-111:2004
22	weight	10kg, E2	2.8 mg	Using E1 Accuracy class

Ratan Singh Rana
Director, IQAS



Scope of Accreditation for Calibration

As per ISO/IEC 17025:2017

Page No. 5 of 6

CAB Name : Tycon Engineering

Address : Plot no. 361,362, samaypur colony near C.I.A office sector-56 Ballabhgarh
Faridabad-121004, Haryana

Certificate No: C-0006

Issue date : 16.01.2025

Validity : 15.01.2029

Amendment date : N/A

Mechanical Calibration

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	CMC(±)	Remarks/Method used
		Accuracy class and Coarser		standard weight 10 kg and Weighing Balance Capacity 10 kg and d: 0.001 g, Substitution Method by ABBA Cycles as per OIML R-111:2004
23	weight	20kg, E2 Accuracy class and Coarser	10 mg	Using E1 Accuracy class standard weight 20 kg and Weighing Balance Capacity 20 kg and d: 0.01 g, Substitution Method by ABBA Cycles as per OIML R-111:2004
24	Weight	50kg, M1 Accuracy class and Coarser	1.0 g	Using F1 Accuracy class standard weight 50 kg and Weighing Balance Capacity 50 kg and d: 1 g, Substitution Method by ABBA Cycles as per OIML R-111:2004
Facility – Site				
01	MECHANICALWEIGHING SCALE AND BALANCE	0 to 5 g, Weighing Balance Class I and Coarser d = 0.0001 mg	0.0080 mg	Using standard Weights E1 Accuracy class 1 mg to 200 g as per OIML R76-1:2006,
02	MECHANICALWEIGHING SCALE AND BALANCE	>5 g to 200 g, Weighing Balance Class I and Coarser d = 0.01 mg	0.040 mg	Using standard Weights E1 Accuracy class 1 mg to 200 g as per OIML R76-1:2006,
03	MECHANICALWEIGHING SCALE AND BALANCE	200g to 300 g, Weighing Balance Class I and Coarser d = 0.1 mg	0.20 mg	Using standard Weights E1 Accuracy class 1 mg to 200 g as per OIML R76-1:2006,

Ratan Singh Rana
Director, IQAS



Scope of Accreditation for Calibration

As per ISO/IEC 17025:2017

Page No. 6 of 6

CAB Name : Tycon Engineering

Address : Plot no. 361,362, samaypur colony near C.I.A office sector-56 Ballabhgarh
Faridabad-121004, Haryana

Certificate No: C-0006

Issue date : 16.01.2025

Validity : 15.01.2029

Amendment date : N/A

Mechanical Calibration

Sr. No.	Parameter/Measurand quantity, Instrument or gauge	Range	CMC(±)	Remarks/Method used
04	MECHANICALWEIGHING SCALE AND BALANCE	>300 g to 1.0 kg Weighing Balance Class I and Coarser d = 0.1 mg	0.20 mg	Using standard Weights E1 Accuracy class 1 mg to 200 g and 500 to 20kg as per OIML R76-1:2006,
05	MECHANICALWEIGHING SCALE AND BALANCE	>1000 g to 10 kg Weighing Balance Class II and Coarser d = 1 mg	3.0 mg	Using standard Weights E1 Accuracy class 1 mg to 200 g and 500 to 20kg as per OIML R76-1:2006,
06	MECHANICALWEIGHING SCALE AND BALANCE	>10kg to 30 kg Weighing Balance Class II and Coarser d = 1 mg	10 mg	Using standard Weights E1 Accuracy class 1 mg to 200 g and 500 to 20kg as per OIML R76-1:2006,
07	MECHANICALWEIGHING SCALE AND BALANCE	>30kg to 50 kg Weighing Balance Class III and Coarser d = 1 mg	2.0 g	Using standard Weights F1 Accuracy class 50kg as per OIML R76-1:2006,
08	MECHANICALWEIGHING SCALE AND BALANCE	>60kg to 150 kg Weighing Balance Class III and Coarser d = 1 mg	20g	Using standard Weights F1 Accuracy class up to 150kg as per OIML R76-1:2006,

Note: CMC is Calibration Measurement Capability expressed as the expanded uncertainty having a coverage factor $k = 2$ at confidence interval to approximately 95%.

Ratan Singh Rana
Director, IQAS